



CERTIFICATE OF ANALYSIS
GemPure Select WFI Quality Water (USP)
(Catalog Number 901-001-022, 901-001-202)

Lot Number: M122098

Date of Manufacture: May2022

Storage Temperature: 2°C – 30°C

For Cell culture, Further Manufacturing or Research use Only. Not for Direct Therapeutic Use.

Test	Methodology	Specification	Analysis
Appearance	Visual Inspection	Clear and colorless liquid	Clear and colorless liquid
Biological Testing			
Endotoxin	USP <85>, EP 2.6.14	≤ 0.25 EU/mL	<0.01 EU/mL
Microbiological Testing			
Sterility	USP <71>, EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Biochemistry			
Total Organic Carbon (Before Packaging)	USP <643>	≤ 500 ppb	17.7 ppb
Total Organic Carbon (Packaged)	USP <643>	≤ 8000 ppb	3670 ppb
Physical Testing			
pH @ 25°C	USP <791>	5.0 – 7.0	5.5
Conductivity @ 25°C (Packaged)	USP <644>, EP 2.2.38	≤ 5 µS/cm	1.28 µS/cm
Conductivity @ 25°C (Before Packaging)	USP <644>, EP 2.2.38	≤ 1.3 µS/cm	0.094 µS/cm

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and it has been confirmed that the testing meets the specifications presented on this Certificate of Analysis.

VANESSA YOUNG

Name

QA ASSOCIATE II

Title

10 JUN 2022

Date