



GEMINI
BIO-PRODUCTS

Certificate of Analysis

Created September 20, 2019

Product: **ITS Media**
Sterile-filtered

Catalog #: 400-145
Lot #: F32T00K

Manufacture Date: Sep 2019
Expiration Date: Sep 2021

TEST	METHODOLOGY	SPECIFICATION	RESULTS
Biological Testing			
Endotoxin	USP<85>	Report	<0.615 EU/mL
Microbiological Testing			
Sterility	USP<71>		
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>	Report	285 mOsm/kg
pH @ RT	USP<791>	Report	7.97

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

Aliu 20 Sep 2019

Created by/ date

Myfan Morgan 20 Sep 2019

Verified by/ date